



Calibration Report

Test # 189755



United States - Main Office

3839 County Road 116

Medina, MN 55340 USA

www.intercompcompany.com

TEL. (763) 476-2531 FAX (763) 476-2613

Intercomp Europe

Manor Farm, Manor Road

Shurlock Row

Berkshire RG10 0PY UK

TEL. 44-118-932-0578 FAX 44-118-932-1034

Tester	DLP		Model #	SW650 RFX BLUETOOTH	
Date	23/May/2022		Scale/Cell ID #		
Indicator			Capacity	1500 lb	
Serial#	0523MS22003		Graduation (d)	1	
Part #	170319		Re-Cal Date	23/May/2023	
Temperature		Humidity		Ack #	PO #

Accuracy: ± 0.1 % of applied load or ± 1 lb - whichever is greater

As Left: In Tolerance

Weight(Lbs.)	Left Front(Lbs.)	Left Rear(Lbs.)	Right Front(Lbs.)	Right Rear(Lbs.)
0	0	0	0	0
200	200	200	200	200
500	500	500	500	500
1000	1000	1000	1000	1000
0	0	0	0	0

Intercomp Company does hereby certify the above listed instrument meets or exceed all published Specifications and has been calibrated using standards whose accuracies are directly traceable to the U.S. National Institute of Standards and Technology.

This document is not to be reproduced except in full without the written permission of Intercomp
Calibration Procedure:

Model #	Test/Trace #	Re-Cal Date
Dead Weight	MN.ST. 400857-1 / 400857-2 / 400833-1 / 400853-1	13/Jul/2024

Tested By:

Title: Calibration Technician

Approve By:

Title: Director