



Calibration Report

Test # 189755



United States - Main Office

3839 County Road 116

Medina, MN 55340 USA

www.intercompcompany.com

TEL. (763) 476-2531 FAX (763) 476-2613

Intercomp Europe

Manor Farm, Manor Road

Shurlock Row

Berkshire RG10 0PY UK

TEL. 44-118-932-0578 FAX 44-118-932-1034

| | | | | | | |
|-------------|-------------|----------|-----------------|---------------------|------|--|
| Tester | DLP | | Model # | SW650 RFX BLUETOOTH | | |
| Date | 23/May/2022 | | Scale/Cell ID # | | | |
| Indicator | | | Capacity | 1500 lb | | |
| Serial# | 0523MS22003 | | Graduation (d) | 1 | | |
| Part # | 170319 | | Re-Cal Date | 23/May/2023 | | |
| Temperature | | Humidity | Ack # | | PO # | |

Accuracy: $\pm 0.1\%$ % of applied load or ± 1 lb - whichever is greater

As Left: In Tolerance

| Weight(Lbs.) | Left Front(Lbs.) | Left Rear(Lbs.) | Right Front(Lbs.) | Right Rear(Lbs.) |
|--------------|------------------|-----------------|-------------------|------------------|
| 0 | 0 | 0 | 0 | 0 |
| 200 | 200 | 200 | 200 | 200 |
| 500 | 500 | 500 | 500 | 500 |
| 1000 | 1000 | 1000 | 1000 | 1000 |
| 0 | 0 | 0 | 0 | 0 |

Intercomp Company does hereby certify the above listed instrument meets or exceed all published Specifications and has been calibrated using standards whose accuracies are directly traceable to the U.S. National Institute of Standards and Technology.

This document is not to be reproduced except in full without the written permission of Intercomp

Calibration Procedure:

| Model # | Test/Trace # | Re-Cal Date |
|-------------|---|-------------|
| Dead Weight | MN.ST. 400857-1 / 400857-2 / 400833-1 / 400853-1 | 13/Jul/2024 |

Tested By: 
Title: Calibration Technician

Approve By: 
Title: Director